

ABSTRACT

u2 In an IVUS system units are located outside or remote from the patient except for the display monitor (12), the catheter interface module (4) and the catheter (3) which are located adjacent the patient together with a control arrangement (13) to enable the said units to be remotely controlled from a position adjacent the patient.

09869637-014502